



E Q U I L I B R I U M

laboratório de controlo de qualidade e de processos



Test Report No.:20217899

Version :1.0

Ambit: Determinations in samples of Medical devices - Face Masks

Definitive Report

Requester: Opharm sp Z o. o.

Address: Pokrzywnica 62 - 99-120 Piatek, Poland

Designation of Sample: SANI-DM-01

SAMPLING

Date:	12/05/2021	Name of Response:	Cliente
Time Collected:	---	V/ Reference:	SANI-DM-01
Methods of Reap:	---	Treatment:	---
Lot:	M001	Expiration:	---
Type:	Medical Device - Masks	Size:	Adult
Composition:	Polypropylene, blue color		

ANALYSIS

Entry Date: 12/05/2021	Period of analysis: 12/05/2021 a 19/05/2021	Ref. Sample: 20217899
------------------------	---	-----------------------

Description	Methods	Unit	Results/Uncertainty	Lim. lei	VMR
Visual Inspection	MI 176 *	-	Suitable	-	-
Bacterial filtration efficiency	EN14683:2019+AC 2019-5.2.2	%	99.3 ±1.8%	TypeI ≥95% TypeII ≥98%	TypeIIR ≥98%
Breathability (Diferencial pressure)	EN14683:2019+AC 2019-5.2.3	Pa/cm2	27,3 ±8.7%	TypeI <40 Type II <40	TypeIIR <60
Microbial cleanliness (Bioburden)	EN ISO 11737-1:2018	UFC/g	3 ±20%	Type I ≤30 Type II ≤30	TypeIIR ≤30
Splash resistance pressure	ISO 22609:2004	---	16.0 /specimens "pass":32, "fail":0	Type I and II not required	Type IIR ≥16.0
Head harness strength	MI 177 *	-	Resists	-	-

APPRECIATION

Read the LimLei and VMR values, the values defined in Table I of EN 14683: 2019.

After comparison, the results obtained in the tests carried out with tabulated limits, the analyzed sample is QUALIFIED - TYPE IIR.

This Test Report refers only to the analyzed samples. Partial reproduction is prohibited.



E Q U I L I B R I U M

laboratório de controlo de qualidade e de processos



Test Report No.:20217899

Version :1.0

Ambit: Determinations in samples of Medical devices - Face Masks

Definitive Report

Requester: Opharm sp Z o. o.

Address: Pokrzywnica 62 - 99-120 Piatek, Poland

Designation of Sample: SANI-DM-01

SAMPLING

Date:	12/05/2021	Name of Response:	Cliente
Time Collected:	---	V/ Reference:	SANI-DM-01
Methods of Reap:	---	Treatment:	---
Lot:	M001	Expiration:	---
Type:	Medical Device - Masks	Size:	Adult
Composition:	Polypropylene, blue color		

ANALYSIS

Entry Date: 12/05/2021	Period of analysis: 12/05/2021 a 19/05/2021	Ref. Sample: 20217899
-------------------------------	--	------------------------------

Description	Methods	Unit	Results/Uncertainty	Lim. lei	VMR
REMARKS					
BACTERIAL FILTRATION EFFICIENCY (Filtration) _EN14683:2019+AC 2019: Condições teste: Temperatura e Humidade relativa // Test conditions: Temperature and Relative humidity: 21±5°C / 85± 5 % Área de cada réplica/ Dimensions of the test specimens: 49cm2 (5 réplicas/ test specimens) Superfície da amostra testada/ Side of the test specimen facing the challenge aerosol: interna / intern Air flow rate: 28.3 l/min. MPS 2.8 Valor médio da contagem total de 2 controlos positivos/Average of the total quantification of 2 positive controls (CFU): 2136 Valor médio da contagem do controle negativo/Average of the quantification of the negative control (CFU): 12 Réplica 1, Réplica 2, Réplica 3, Réplica 4, Réplica 5 (%) // Test specimen 1 ,Test specimen 2,Test specimen 3,Test specimen 4,Test specimen 5 (%) 99.3 99.3 99.2 99.3 99.2					
DIFFERENCIAL PRESSURE (Breathability) _EN14683:2019+AC 2019: Condições teste: Temperatura e Humidade relativa // Test conditions: Temperature and Relative humidity: 21±5°C / 85± 5 % Área de cada réplica/ Dimensions of the test specimens: 4.9cm2 (5 réplicas/ 5 test specimens) Número de áreas por réplica// Number of áreas per specimen: 5 áreas por réplica (A,B,C,D,E)// 5 áreas for test specimen (A,B,C,D,E) Superfície da amostra testada: Teste realizado com a direcção do fluxo do interior para o exterior. Localização lateral e central. // Number and general location of the areas of the mask the differential measurements were taken: Test performed with the direction of flow from the inside to the outside. Side and central location. Air flow rate: 8L/min Réplica 1 (área A,B,C,D,E), Réplica 2 (área A,B,C,D,E), Réplica 3 (área A,B,C,D,E), Réplica 4 (área A,B,C,D,E) Réplica 5 (área A,B,C,D,E) //Test specimen 1 (área A,B,C,D,E), Test specimen 2 (área A,B,C,D,E),Test specimen 3 (área A,B,C,D,E), Test specimen 4 (área A,B,C,D,E), Test specimen 5 (área A,B,C,D,E) Unidades (Pa/cm2) 29,6 27,6 26,5 26,5 26,5					
SPLASH RESISTANCE PRESSURE _ISO 22609:2004:					

This Test Report refers only to the analyzed samples. Partial reproduction is prohibited.



E Q U I L I B R I U M

laboratório de controlo de qualidade e de processos



Test Report No.:20217899

Version :1.0

Ambit: Determinations in samples of Medical devices - Face Masks

Definitive Report

Requester: Opharm sp Z o. o.

Address: Pokrzywnica 62 - 99-120 Piatek, Poland

Designation of Sample: SANI-DM-01

SAMPLING

Date:	12/05/2021	Name of Response:	Cliente
Time Collected:	---	V/ Reference:	SANI-DM-01
Methods of Reap:	---	Treatment:	---
Lot:	M001	Expiration:	---
Type:	Medical Device - Masks	Size:	Adult
Composition:	Polypropylene, blue color		

ANALYSIS

Entry Date: 12/05/2021	Period of analysis: 12/05/2021 a 19/05/2021	Ref. Sample: 20217899
------------------------	---	-----------------------

Description	Methods	Unit	Results/Uncertainty	Lim. lei	VMR
Condições teste: Amostra pré-condicionadas pelo menos 4 horas a Temperatura e Humidade relativa de $21 \pm 5^\circ\text{C}$ / $85 \pm 5\%$ // Test conditions: Samples pre-conditioned for at least 4 hours at Temperature and Relative humidity: $21 \pm 5^\circ\text{C}$ / $85 \pm 5\%$ Amostra expostas a um esguicho de 2 mL sangue sintético a diferentes pressões (baixa: 10.6 KPa; média 16.0 KPa; alta 21.3 KPa), tendo como área alvo o centro da máscara // Samples exposed to a jet of 2 mL synthetic blood at pressure (low: 10.6 KPa; medium: 16.0 KPa; high: 21.3 KPa) aimed at the centre of the mask. Teste realizado à temperatura do laboratório de 21°C e 45% humidade relativa, após 60 segundos da remoção da câmara de condicionamento // Test performed at laboratory temperature of 21°C and 45% relative humidity, within 60 seconds after the mask was removed from the conditioning chamber Observação após 10+1 segundo da penetração do sangue do lado oposto da máscara // Observation after 10+1 second of blood penetration on the opposite side of the mask. Sangue sintético segundo o anexo B da ISO 22609: 2004 com tensão superficial de $42 \pm 2 \text{ mN/m}$, lote L202015 // Synthetic blood according to Annex B of ISO 22609: 2004 with surface tension of $42 \pm 2 \text{ mN/m}$, batch L202015 Número e localização geral das áreas: 32 / centro da amostra (Se passar no mínimo um teste de pressão média para 29 das 32 amostras, correspondendo a um AQL 4%, conforme a EN 14683: 2019 máscara tipo IIR) // Number and General location of the áreas: 32 test specimen / center (pass at least Medium pressure test for 29 out of 32 samples as minimum, corresponding to AQL 4%, according EN 14683: 2019 mask Type IIR) Quantidade de amostras: 32 // 32 test specimen Localização alvo de cada réplica // General location of the áreas: centro da amostra // center AQL 4%: passando no mínimo 29 das 32 amostras analisadas para a pressão testada, conforme a EN 14683: 2019 máscara tipo IIR) // pass at least Medium pressure test for 29 out of 32 samples as minimum, corresponding to AQL 4%, according EN 14683: 2019 mask Type IIR. Resultados: Pressão baixa // Low pressure (10.6 KPa) _ amostras/specimen "pass", _ amostras/specimen "fail" Pressão média // médium pressure (16.0 KPa) 32 amostras/specimen "pass", 0 amostras/specimen "fail"					
Amost1	--	Pass			
Amost2	--	Pass			
Amost3	--	Pass			
Amost4	--	Pass			
Amost5	--	Pass			
Amost6	--	Pass			
Amost7	--	Pass			
Amost8	--	Pass			
Amost9	--	Pass			
Amost10	--	Pass			
Amost11	--	Pass			
Amost12	--	Pass			

This Test Report refers only to the analyzed samples. Partial reproduction is prohibited.



E Q U I L I B R I U M

laboratório de controlo de qualidade e de processos



Test Report No.:20217899

Version :1.0

Ambit: Determinations in samples of Medical devices - Face Masks

Definitive Report

Requester: Opharm sp Z o. o.

Address: Pokrzywnica 62 - 99-120 Piatek, Poland

Designation of Sample: SANI-DM-01

SAMPLING

Date:	12/05/2021	Name of Response:	Cliente
Time Collected:	---	V/ Reference:	SANI-DM-01
Methods of Reap:	---	Treatment:	---
Lot:	M001	Expiration:	---
Type:	Medical Device - Masks	Size:	Adult
Composition:	Polypropylene, blue color		

ANALYSIS

Entry Date: 12/05/2021	Period of analysis: 12/05/2021 a 19/05/2021	Ref. Sample: 20217899
------------------------	---	-----------------------

Description	Methods	Unit	Results/Uncertainty	Lim. lei	VMR
Amo13 -- Pass					
Amo14 -- Pass					
Amo15 -- Pass					
Amo16 -- Pass					
Amo17 -- Pass					
Amo18 -- Pass					
Amo19 -- Pass					
Amo20 -- Pass					
Amo21 -- Pass					
Amo22 -- Pass					
Amo23 -- Pass					
Amo24 -- Pass					
Amo25 -- Pass					
Amo26 -- Pass					
Amo27 -- Pass					
Amo28 -- Pass					
Amo29 -- Pass					
Amo30 -- Pass					
Amo31 -- Pass					
Amo32 -- Pass					
MICROBIAL CLEANLINESS (Bioburden) _ EN ISO 11737-1:2018:					
5 min shaker at 250rpm					
Área of each test specimen: 5 test specimens					
Mic30°C (3 days), Molds and yeasts 25°C (7 days)					
Medida de Incerteza Expandida a um nível de confiança de 95%, k=2/ The expanded uncertainty at a confidence level of 95%, K=2: 20%					
3					
2					
4					
2					
2					
RESISTÊNCIA AO ELÁSTICO/ Head harness strength :					
Verificação da Resistência a 5 ciclos de colocação e remoção, em 3 sujeitos de ensaio, com diferentes morfologias.					

This Test Report refers only to the analyzed samples. Partial reproduction is prohibited.



E Q U I L I B R I U M

laboratório de controlo de qualidade e de processos



Test Report No.:20217899

Version :1.0

Ambit: Determinations in samples of Medical devices - Face Masks

Definitive Report

Requester: Opharm sp Z o. o.

Address: Pokrzywnica 62 - 99-120 Piatek, Poland

Designation of Sample: SANI-DM-01

SAMPLING

Date:	12/05/2021	Name of Response:	Cliente
Time Collected:	---	V/ Reference:	SANI-DM-01
Methods of Reap:	---	Treatment:	---
Lot:	M001	Expiration:	---
Type:	Medical Device - Masks	Size:	Adult
Composition:	Polypropylene, blue color		

ANALYSIS

Entry Date: 12/05/2021	Period of analysis: 12/05/2021 a 19/05/2021	Ref. Sample: 20217899
-------------------------------	--	------------------------------

Description	Methods	Unit	Results/Uncertainty	Lim. lei	VMR
// Verification of resistance to 5 cycles of placement and removal, in 3 test subjects, with different morphologies.					
Inspeção Visual/ Visual Inspection: - Ajuste do clipe nasal // Nasal clip adjustment - Rasgos // Rips, - Deformações // Deformations, - Desgaste // Wear, - Cobertura Facial // Facial Coverage					

This Test Report refers only to the analyzed samples. Partial reproduction is prohibited.



E Q U I L I B R I U M

laboratório de controlo de qualidade e de processos



Test Report No.:20217899

Version :1.0

Ambit: Determinations in samples of Medical devices - Face Masks

Definitive Report

Requester: Opharm sp Z o. o.

Address: Pokrzywnica 62 - 99-120 Piatek, Poland

Designation of Sample: SANI-DM-01


SAMPLING

Date:	12/05/2021	Name of Response:	Cliente
Time Collected:	---	V/ Reference:	SANI-DM-01
Methods of Reap:	---	Treatment:	---
Lot:	M001	Expiration:	---
Type:	Medical Device - Masks	Size:	Adult
Composition:	Polypropylene, blue color		

ANALYSIS

Entry Date: 12/05/2021	Period of analysis: 12/05/2021 a 19/05/2021	Ref. Sample: 20217899
-------------------------------	--	------------------------------

Description	Methods	Unit	Results/Uncertainty	Lim. lei	VMR
-------------	---------	------	---------------------	----------	-----



This Test Report refers only to the analyzed samples. Partial reproduction is prohibited.



E Q U I L I B R I U M

laboratório de controlo de qualidade e de processos



Test Report No.:20217899

Version :1.0

Ambit: Determinations in samples of Medical devices - Face Masks

Definitive Report

Requester: Opharm sp Z o. o.

Address: Pokrzywnica 62 - 99-120 Piatek, Poland

Designation of Sample: SANI-DM-01

SAMPLING

Date:	12/05/2021	Name of Response:	Cliente
Time Collected:	---	V/ Reference:	SANI-DM-01
Methods of Reap:	---	Treatment:	---
Lot:	M001	Expiration:	---
Type:	Medical Device - Masks	Size:	Adult
Composition:	Polypropylene, blue color		

ANALYSIS

Entry Date: 12/05/2021	Period of analysis: 12/05/2021 a 19/05/2021	Ref. Sample: 20217899
------------------------	---	-----------------------

Description	Methods	Unit	Results/Uncertainty	Lim. lei	VMR
-------------	---------	------	---------------------	----------	-----

The indicated test with (*) isn't included on the ambit of accreditation of the laboratory

The Sampling is not included on the scope of accreditation.

NP: Norma Portuguesa; SMEWW: Standart Methods for the Examination of Water and Wastewater; ISO: Internacional Standard Organization; AFNOR: Association Francaise de Normalisation; LAE: L'Analyse des Eaux, Rodier 8e Édition; EN: Norma Europeia; DIN: Deutsches Institut für Normung; EPA: Environmental Protection Agency; ASTM: American Society for Testing and Materials; MI: Método Interno; N/A: Não aplicável; LQ: Limite de Quantificação; LD: Limite de Detecção; UFC: Unidades formadoras de colónias; VMR: Valor máximo recomendado; VMA: Valor máximo admissível; VP: Valor Paramétrico; Nos resultados obtidos por cálculo com base em resultados individuais, serão contabilizadas as parcelas quantificáveis desprezando as parcelas <LQ. Se todas as parcelas forem <LQ, o valor emitido será o LQ do método. O resultado pode ser apresentado com Medida de Incerteza Expandida, do Ensaio (não contemplando a da amostragem), a um nível de confiança de 95%, k=2, quando esta aparece associada ao resultado. O laboratório não contabiliza a incerteza no método na declaração de conformidade. A apresentação do resultado como $\leq$ ao LQ corresponde à faixa de guarda associada à incerteza do método. Os dados fornecidos pelo cliente são da sua responsabilidade (descritos no campo: "Colheitas" e "Designação Amostras"). Os resultados aplicam-se à amostra conforme recepcionada/Results apply to the sample as received.

Emission

Matosinhos, 21 de maio de 2021

Laboratory Director

Maria Cristina Antão

Maria Cristina Antão, Dra.

(This Test Report was digitally signed)

This Test Report refers only to the analyzed samples. Partial reproduction is prohibited.